

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
 Data File : VD073364.D
 Acq On : 25 May 2022 16:53
 Operator : VA/SY
 Sample : N3027-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-03-052422

Quant Time: May 26 02:25:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	19068	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	30894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	37228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	16256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	17952	85.190	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 170.380%#	
35) Dibromofluoromethane	7.908	113	16046	78.711	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 157.420%#	
50) Toluene-d8	10.332	98	40260	53.575	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.140%	
62) 4-Bromofluorobenzene	12.620	95	15593	59.126	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 118.260%	
Target Compounds						Qvalue
3) Chloromethane	2.208	50	807	3.746	ug/l	89
16) Acetone	4.150	43	763	20.904	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
Data File : VD073364.D
Acq On : 25 May 2022 16:53
Operator : VA/SY
Sample : N3027-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-03-052422

Quant Time: May 26 02:25:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

