

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
 Data File : VD073366.D
 Acq On : 25 May 2022 17:51
 Operator : VA/SY
 Sample : N2949-17RE
 Misc : 5.76G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P044-SS001-1824-01

Quant Time: May 26 02:25:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	101555	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	165387	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	194193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	87538	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	72348	64.463	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.920%	
35) Dibromofluoromethane	7.914	113	66848	61.254	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 122.500%	
50) Toluene-d8	10.332	98	206712	51.384	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.760%	
62) 4-Bromofluorobenzene	12.620	95	81163	57.489	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 114.980%	
Target Compounds						
16) Acetone	4.162	43	2479	12.752	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
Data File : VD073366.D
Acq On : 25 May 2022 17:51
Operator : VA/SY
Sample : N2949-17RE
Misc : 5.76G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P044-SS001-1824-01

Quant Time: May 26 02:25:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

