

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
 Data File : VD073370.D
 Acq On : 25 May 2022 19:46
 Operator : VA/SY
 Sample : N2886-14
 Misc : 4.45G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P021-SS005-0006-01

Quant Time: May 26 04:07:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	287954	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	512571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	526421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	229776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	155945	49.004	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.000%
35) Dibromofluoromethane	7.908	113	173593	51.324	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.640%
50) Toluene-d8	10.332	98	622908	49.961	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.920%
62) 4-Bromofluorobenzene	12.620	95	220449	50.382	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.760%
Target Compounds						
11) Tert butyl alcohol	5.167	59	18495	37.459	ug/l	Qvalue # 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
Data File : VD073370.D
Acq On : 25 May 2022 19:46
Operator : VA/SY
Sample : N2886-14
Misc : 4.45G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P021-SS005-0006-01

Quant Time: May 26 04:07:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

