

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
 Data File : VD073373.D
 Acq On : 25 May 2022 21:13
 Operator : VA/SY
 Sample : N2886-17
 Misc : 4.50G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P030-SS001-0006-01

Quant Time: May 26 02:27:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	285961	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	498484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	523418	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	232686	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	152507	48.258	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.520%
35) Dibromofluoromethane	7.909	113	172718	52.509	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.020%
50) Toluene-d8	10.332	98	619617	51.101	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.200%
62) 4-Bromofluorobenzene	12.620	95	217186	51.039	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.080%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.985	85	4998	1.749	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
Data File : VD073373.D
Acq On : 25 May 2022 21:13
Operator : VA/SY
Sample : N2886-17
Misc : 4.50G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P030-SS001-0006-01

Quant Time: May 26 02:27:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

