

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
 Data File : VD073376.D
 Acq On : 25 May 2022 22:39
 Operator : VA/SY
 Sample : N2886-20
 Misc : 5.53G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P030-SS001-1824-01

Quant Time: May 26 02:27:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	274734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	478078	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	517669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	235202	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	157325	51.816	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.640%	
35) Dibromofluoromethane	7.908	113	164512	52.149	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.300%	
50) Toluene-d8	10.332	98	607197	52.214	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.420%	
62) 4-Bromofluorobenzene	12.620	95	220542	54.040	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.080%	
Target Compounds						Qvalue
16) Acetone	4.156	43	7490	14.242	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
Data File : VD073376.D
Acq On : 25 May 2022 22:39
Operator : VA/SY
Sample : N2886-20
Misc : 5.53G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P030-SS001-1824-01

Quant Time: May 26 02:27:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

