

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
 Data File : VD073377.D
 Acq On : 25 May 2022 23:07
 Operator : VA/SY
 Sample : N2886-01
 Misc : 5.46G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P014-SS002-2430-01

Quant Time: May 26 02:27:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	287136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	502115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	534777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	246759	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	152430	48.036	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.080%
35) Dibromofluoromethane	7.909	113	171425	51.739	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.480%
50) Toluene-d8	10.332	98	622113	50.936	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.880%
62) 4-Bromofluorobenzene	12.620	95	228719	53.361	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052522\
Data File : VD073377.D
Acq On : 25 May 2022 23:07
Operator : VA/SY
Sample : N2886-01
Misc : 5.46G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P014-SS002-2430-01

Quant Time: May 26 02:27:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

