

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
 Data File : VD076236.D
 Acq On : 25 May 2023 16:39
 Operator : KP/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: May 26 02:00:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	164249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	311868	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	283908	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	125225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	88257	55.598	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	111.200%	
35) Dibromofluoromethane	7.805	113	105218	48.809	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	97.620%	
50) Toluene-d8	10.269	98	357280	42.516	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	85.040%	
62) 4-Bromofluorobenzene	12.575	95	95360	40.851	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	81.700%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.799	84	55885	15.939	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
Data File : VD076236.D
Acq On : 25 May 2023 16:39
Operator : KP/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: May 26 02:00:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

