

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
 Data File : VD076243.D
 Acq On : 25 May 2023 20:08
 Operator : KP/SY
 Sample : 02938-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-04-052423

Quant Time: May 26 02:01:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	4837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	10548	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	11094	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	4858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	6842	146.358	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	292.720%#	
35) Dibromofluoromethane	7.804	113	6106	83.747	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	167.500%#	
50) Toluene-d8	10.269	98	14011	49.296	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	98.600%	
62) 4-Bromofluorobenzene	12.569	95	4608	58.364	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	116.720%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.805	84	1895	19.988	ug/l	# 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
Data File : VD076243.D
Acq On : 25 May 2023 20:08
Operator : KP/SY
Sample : 02938-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-04-052423

Quant Time: May 26 02:01:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

