

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
 Data File : VD076244.D
 Acq On : 25 May 2023 20:38
 Operator : KP/SY
 Sample : 02940-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RBR-251671

Quant Time: May 26 02:01:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	147906	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	309105	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	280559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	104908	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	106853	74.750	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 149.500%	
35) Dibromofluoromethane	7.805	113	135250	63.302	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 126.600%	
50) Toluene-d8	10.269	98	355226	42.649	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 85.300%	
62) 4-Bromofluorobenzene	12.575	95	117770	50.902	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.800%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.799	84	46815	14.086	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
Data File : VD076244.D
Acq On : 25 May 2023 20:38
Operator : KP/SY
Sample : 02940-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RBR-251671

Quant Time: May 26 02:01:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

