

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
 Data File : VD076246.D
 Acq On : 25 May 2023 21:38
 Operator : KP/SY
 Sample : 02919-01
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 IMPACTED-STONE

Quant Time: May 26 02:01:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	207557	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	437809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	339088	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	100803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	129848	64.730	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.460%
35) Dibromofluoromethane	7.805	113	177153	58.539	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.080%
50) Toluene-d8	10.269	98	457847	38.811	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	77.620%
62) 4-Bromofluorobenzene	12.575	95	127659	38.956	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.920%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.793	84	58163	11.265	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
Data File : VD076246.D
Acq On : 25 May 2023 21:38
Operator : KP/SY
Sample : 02919-01
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
IMPACTED-STONE

Quant Time: May 26 02:01:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

