

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
 Data File : VD076247.D
 Acq On : 25 May 2023 22:08
 Operator : KP/SY
 Sample : 02923-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-03-052423

Quant Time: May 26 02:01:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	1225	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	1038	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	192	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	710	394.962	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	789.920%#
35) Dibromofluoromethane	7.804	113	325	38.382	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.760%
50) Toluene-d8	10.263	98	1416	42.898	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	85.800%
62) 4-Bromofluorobenzene	12.575	95	215	23.448	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	46.900%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.798	84	55	12.465	ug/l #	4
76) 1,2,3-Trichloropropane	12.569	75	61	36.306	ug/l #	100
81) trans-1,4-Dichloro-2-b...	12.569	75	61	76.164	ug/l #	5

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052523\
Data File : VD076247.D
Acq On : 25 May 2023 22:08
Operator : KP/SY
Sample : 02923-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-03-052423

Quant Time: May 26 02:01:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

