

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052622\
 Data File : VD073385.D
 Acq On : 26 May 2022 13:43
 Operator : VA/SY
 Sample : N3001-01
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BU-03-052322

Quant Time: May 26 16:58:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	200705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	346478	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	377124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	166480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	131629	59.344	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 118.680%	
35) Dibromofluoromethane	7.908	113	129710	56.734	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.460%	
50) Toluene-d8	10.332	98	433908	51.485	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.980%	
62) 4-Bromofluorobenzene	12.620	95	159789	54.025	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.060%	
Target Compounds						
16) Acetone	4.150	43	16347	42.549	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052622\
Data File : VD073385.D
Acq On : 26 May 2022 13:43
Operator : VA/SY
Sample : N3001-01
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BU-03-052322

Quant Time: May 26 16:58:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

