

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052622\
 Data File : VD073386.D
 Acq On : 26 May 2022 14:12
 Operator : VA/SY
 Sample : N2886-13
 Misc : 5.23G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P021-SS004-0006-01

Quant Time: May 27 01:31:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	301498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	527969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	544956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	253324	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	149468	44.859	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.720%
35) Dibromofluoromethane	7.914	113	176749	50.733	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.460%
50) Toluene-d8	10.332	98	660321	51.417	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.840%
62) 4-Bromofluorobenzene	12.620	95	243864	54.108	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.220%
Target Compounds						
20) Methylene Chloride	4.961	84	46287	9.402	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052622\
Data File : VD073386.D
Acq On : 26 May 2022 14:12
Operator : VA/SY
Sample : N2886-13
Misc : 5.23G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P021-SS004-0006-01

Quant Time: May 27 01:31:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

