

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052622\
 Data File : VD073389.D
 Acq On : 26 May 2022 15:37
 Operator : VA/SY
 Sample : N3041-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 2147

Quant Time: May 27 01:31:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	253694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	451822	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	456419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	166926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	143545	51.199	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.400%	
35) Dibromofluoromethane	7.908	113	158735	53.241	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.480%	
50) Toluene-d8	10.332	98	556889	50.671	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.340%	
62) 4-Bromofluorobenzene	12.620	95	178914	46.388	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.780%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.955	84	33802	7.587	ug/l	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052622\
Data File : VD073389.D
Acq On : 26 May 2022 15:37
Operator : VA/SY
Sample : N3041-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
2147

Quant Time: May 27 01:31:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

