

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052623\
 Data File : VD076257.D
 Acq On : 26 May 2023 14:12
 Operator : KP/SY
 Sample : 02913-01RE
 Misc : 6.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-70A

Quant Time: May 26 23:16:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	263904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	525489	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	464296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	194194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	147244	57.730	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.460%
35) Dibromofluoromethane	7.805	113	204370	56.265	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.540%
50) Toluene-d8	10.269	98	611184	43.164	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	86.320%
62) 4-Bromofluorobenzene	12.575	95	208971	53.128	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.260%
Target Compounds						
16) Acetone	4.022	43	9528	18.776	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052623\
Data File : VD076257.D
Acq On : 26 May 2023 14:12
Operator : KP/SY
Sample : 02913-01RE
Misc : 6.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-70A

Quant Time: May 26 23:16:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

