

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052623\
 Data File : VD076259.D
 Acq On : 26 May 2023 15:10
 Operator : KP/SY
 Sample : 02944-17RE
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-VPB-RW11-GW-418-420RE

Quant Time: May 26 16:05:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	190031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	371936	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	290420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	79635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	118933	64.757	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery	= 129.520%		
35) Dibromofluoromethane	7.804	113	157294	61.183	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery	= 122.360%		
50) Toluene-d8	10.269	98	433373	43.242	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery	= 86.480%		
62) 4-Bromofluorobenzene	12.575	95	115769	41.584	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery	= 83.160%		
Target Compounds						Qvalue
16) Acetone	4.016	43	3189	8.727	ug/l #	78
20) Methylene Chloride	4.799	84	46886	8.675	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052623\
Data File : VD076259.D
Acq On : 26 May 2023 15:10
Operator : KP/SY
Sample : 02944-17RE
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-VPB-RW11-GW-418-420RE

Quant Time: May 26 16:05:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

