

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052623\
 Data File : VD076265.D
 Acq On : 26 May 2023 18:39
 Operator : KP/SY
 Sample : 02888-04
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FS-SB-20(ELEV-13-11)

Quant Time: May 26 23:17:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	240596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	481396	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	408132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	148297	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	137009	58.921	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.840%
35) Dibromofluoromethane	7.804	113	192033	57.711	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.420%
50) Toluene-d8	10.269	98	550398	42.432	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.860%
62) 4-Bromofluorobenzene	12.575	95	179561	49.833	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.660%
Target Compounds						
52) Toluene	10.334	92	24873	2.756	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052623\
Data File : VD076265.D
Acq On : 26 May 2023 18:39
Operator : KP/SY
Sample : 02888-04
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FS-SB-20(ELEV-13-11)

Quant Time: May 26 23:17:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

