



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052623\
Data File : VD076249.D
Acq On : 26 May 2023 09:35
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Title : SW846 8260
Last Update : Tue May 16 10:08:53 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.6	5874	PASS
75	95	30	60	51.9	16410	PASS
95	95	100	100	100.0	31613	PASS
96	95	5	9	6.3	1983	PASS
173	174	0.00	2	1.4	446	PASS
174	95	50	100	98.0	30971	PASS
175	174	5	9	7.5	2329	PASS
176	174	95	101	95.5	29589	PASS
177	176	5	9	6.6	1960	PASS