

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053023\
 Data File : VD076275.D
 Acq On : 30 May 2023 15:30
 Operator : KP/SY
 Sample : 02971-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB16

Quant Time: May 31 02:20:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	7295	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.769	114	13871	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	12963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	7155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	6868	97.412	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 194.820%#	
35) Dibromofluoromethane	7.804	113	6503	67.825	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 135.660%	
50) Toluene-d8	10.269	98	11331	30.316	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 60.640%	
62) 4-Bromofluorobenzene	12.569	95	4715	45.413	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 90.820%	
Target Compounds						
					Qvalue	
5) Bromomethane	2.699	94	1566	16.393	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053023\
Data File : VD076275.D
Acq On : 30 May 2023 15:30
Operator : KP/SY
Sample : 02971-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB16

Quant Time: May 31 02:20:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

