

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053023\
 Data File : VD076277.D
 Acq On : 30 May 2023 16:30
 Operator : KP/SY
 Sample : 02971-05
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB12

Quant Time: May 31 02:20:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	4584	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	11569	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	16106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	7296	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	7377	166.512	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	333.020%#
35) Dibromofluoromethane	7.810	113	7335	91.725	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	183.460%#
50) Toluene-d8	10.263	98	16595	53.235	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.460%
62) 4-Bromofluorobenzene	12.575	95	5649	65.235	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	130.460%
Target Compounds						
						Qvalue
5) Bromomethane	2.699	94	1149	19.141	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053023\
Data File : VD076277.D
Acq On : 30 May 2023 16:30
Operator : KP/SY
Sample : 02971-05
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB12

Quant Time: May 31 02:20:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

