

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053023\
 Data File : VD076279.D
 Acq On : 30 May 2023 17:30
 Operator : KP/SY
 Sample : 02965-01
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ETGI-338

Quant Time: May 31 02:21:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	230573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	450065	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	422127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	181363	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	123678	55.500	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.000%
35) Dibromofluoromethane	7.804	113	164154	52.767	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.540%
50) Toluene-d8	10.269	98	439027	36.202	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	72.400%
62) 4-Bromofluorobenzene	12.575	95	164985	48.975	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.940%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053023\
Data File : VD076279.D
Acq On : 30 May 2023 17:30
Operator : KP/SY
Sample : 02965-01
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ETGI-338

Quant Time: May 31 02:21:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD076279.D
Acq: 30 May 2023 17:30




Instrument :
MSVOA_D
ClientSampleId :
ETGI-338


Tgt Ion:168 Resp: 230573
Ion Ratio Lower Upper
168 100
99 45.7 43.8 65.6



#33
1,2-Dichloroethane-d4
Concen: 55.500 ug/l
RT: 8.222 min Scan# 1128
Delta R.T. -0.012 min
Lab File: VD076279.D
Acq: 30 May 2023 17:30

Tgt Ion: 65 Resp: 123678
Ion Ratio Lower Upper
65 100
67 54.9 0.0 102.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.775 min Scan# 114
Delta R.T. -0.000 min
Lab File: VD076279.D
Acq: 30 May 2023 17:30




Instrument :
MSVOA_D
ClientSampleId :
ETGI-338


Tgt Ion:114 Resp: 450065
Ion Ratio Lower Upper
114 100
63 16.1 0.0 41.6
88 13.6 0.0 30.6



#35
Dibromofluoromethane
Concen: 52.767 ug/l
RT: 7.804 min Scan# 1057
Delta R.T. -0.001 min
Lab File: VD076279.D
Acq: 30 May 2023 17:30

Tgt Ion:113 Resp: 164154
Ion Ratio Lower Upper
113 100
111 103.2 81.4 122.0
192 19.9 15.2 22.8









#50
Toluene-d8
Concen: 36.202 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VD076279.D
Acq: 30 May 2023 17:30




Instrument :
MSVOA_D
ClientSampleId :
ETGI-338


Tgt Ion: 98 Resp: 439027
Ion Ratio Lower Upper
98 100
100 62.8 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 48.975 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD076279.D
Acq: 30 May 2023 17:30

Tgt Ion: 95 Resp: 164985
Ion Ratio Lower Upper
95 100
174 95.5 0.0 164.0
176 90.9 0.0 156.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. -0.000 min
Lab File: VD076279.D
Acq: 30 May 2023 17:30

Instrument :
MSVOA_D
ClientSampleId :
ETGI-338

Tgt Ion:117 Resp: 422127
Ion Ratio Lower Upper
117 100
82 51.5 44.0 66.0
119 31.3 25.1 37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.522 min Scan# 2029
Delta R.T. -0.000 min
Lab File: VD076279.D
Acq: 30 May 2023 17:30

Tgt Ion:152 Resp: 181363
Ion Ratio Lower Upper
152 100
115 52.7 30.1 90.3
150 156.6 0.0 364.8

