


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060222\
 Data File : VD073460.D
 Acq On : 02 Jun 2022 09:26
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Title : SW846 8260
 Last Update : Thu May 19 15:13:42 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	33744	PASS
75	95	30	60	50.8	98619	PASS
95	95	100	100	100.0	194091	PASS
96	95	5	9	6.8	13264	PASS
173	174	0.00	2	1.5	2291	PASS
174	95	50	100	77.8	150992	PASS
175	174	5	9	7.3	11016	PASS
176	174	95	101	95.8	144605	PASS
177	176	5	9	6.7	9679	PASS