

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060723\
 Data File : VD076362.D
 Acq On : 07 Jun 2023 19:19
 Operator : KP/SY
 Sample : 03042-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1

Quant Time: Jun 08 01:19:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	66627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	135633	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	146747	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	66090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	65629	109.936	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 219.880%#	
35) Dibromofluoromethane	7.805	113	66894	77.677	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 155.360%#	
50) Toluene-d8	10.269	98	141489	41.265	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 82.540%	
62) 4-Bromofluorobenzene	12.575	95	49917	55.011	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 110.020%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.216	105	6747	1.936	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060723\
Data File : VD076362.D
Acq On : 07 Jun 2023 19:19
Operator : KP/SY
Sample : 03042-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1

Quant Time: Jun 08 01:19:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

