

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060723\
 Data File : VD076364.D
 Acq On : 07 Jun 2023 20:18
 Operator : KP/SY
 Sample : 03047-02RE
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB32BRE

Quant Time: Jun 08 01:19:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	7998	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	16371	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	20155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	9206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	10889	151.950	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 303.900%#	
35) Dibromofluoromethane	7.804	113	10051	96.695	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 193.400%#	
50) Toluene-d8	10.269	98	18541	44.800	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 89.600%	
62) 4-Bromofluorobenzene	12.575	95	7553	68.962	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 137.920%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060723\
Data File : VD076364.D
Acq On : 07 Jun 2023 20:18
Operator : KP/SY
Sample : 03047-02RE
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB32BRE

Quant Time: Jun 08 01:19:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

