

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060723\
 Data File : VD076357.D
 Acq On : 07 Jun 2023 16:49
 Operator : KP/SY
 Sample : 03032-01
 Misc : 5.42G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-20-0A

Quant Time: Jun 08 01:18:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.875	168	210848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	399747	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	384827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	167261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	107933	57.132	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.260%
35) Dibromofluoromethane	7.805	113	146864	57.863	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.720%
50) Toluene-d8	10.269	98	424733	42.030	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.060%
62) 4-Bromofluorobenzene	12.575	95	130660	48.857	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.720%
Target Compounds						
16) Acetone	4.034	43	4682	13.531	ug/l	Qvalue 96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060723\
Data File : VD076357.D
Acq On : 07 Jun 2023 16:49
Operator : KP/SY
Sample : 03032-01
Misc : 5.42G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-20-0A


Quant Time: Jun 08 01:18:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD076357.D
Acq: 07 Jun 2023 16:49

Instrument :
MSVOA_D
ClientSampleId :
TP-20-0A







Tgt Ion:168 Resp: 210848
Ion Ratio Lower Upper
168 100
99 45.4 36.7 55.1



#16
Acetone
Concen: 13.531 ug/l
RT: 4.034 min Scan# 416
Delta R.T. 0.012 min
Lab File: VD076357.D
Acq: 07 Jun 2023 16:49

Tgt Ion: 43 Resp: 4682
Ion Ratio Lower Upper
43 100
58 46.1 34.7 52.1





#33

1,2-Dichloroethane-d4

Concen: 57.132 ug/l

RT: 8.228 min Scan# 1129

Delta R.T. -0.000 min

Lab File: VD076357.D

Acq: 07 Jun 2023 16:49

Instrument :

MSVOA_D

ClientSampleId :





TP-20-0A

Tgt Ion: 65 Resp: 107933

Ion Ratio Lower Upper

65 100

67 55.0 0.0 115.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. -0.000 min

Lab File: VD076357.D

Acq: 07 Jun 2023 16:49




Tgt Ion: 114 Resp: 399747


Ion Ratio Lower Upper

114 100

63 17.2 0.0 33.4

88 13.6 0.0 26.2









#35
Dibromofluoromethane
Concen: 57.863 ug/l
RT: 7.805 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VD076357.D
Acq: 07 Jun 2023 16:49




Instrument :
MSVOA_D
ClientSampleId :
TP-20-0A


Tgt Ion:113 Resp: 146864
Ion Ratio Lower Upper
113 100
111 103.3 84.7 127.1
192 20.8 16.8 25.2



#50
Toluene-d8
Concen: 42.030 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VD076357.D
Acq: 07 Jun 2023 16:49

Tgt Ion: 98 Resp: 424733
Ion Ratio Lower Upper
98 100
100 64.4 50.3 75.5









#62
4-Bromofluorobenzene
Concen: 48.857 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD076357.D
Acq: 07 Jun 2023 16:49




Instrument :
MSVOA_D
ClientSampleId :
TP-20-0A


Tgt Ion: 95 Resp: 130660
Ion Ratio Lower Upper
95 100
174 98.5 0.0 201.8
176 95.0 0.0 191.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. -0.000 min
Lab File: VD076357.D
Acq: 07 Jun 2023 16:49


Tgt Ion: 117 Resp: 384827
Ion Ratio Lower Upper
117 100
82 45.9 38.3 57.5
119 31.8 25.0 37.4





#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.522 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: VD076357.D
 Acq: 07 Jun 2023 16:49

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-20-0A



Tgt Ion:152 Resp: 167261
 Ion Ratio Lower Upper
 152 100
 115 50.3 27.1 81.3
 150 157.7 0.0 353.6

