



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060723\
Data File : VD076350.D
Acq On : 07 Jun 2023 11:51
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Title : SW846 8260
Last Update : Tue Jun 06 02:52:23 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	7778	PASS
75	95	30	60	49.1	23893	PASS
95	95	100	100	100.0	48691	PASS
96	95	5	9	6.4	3092	PASS
173	174	0.00	2	1.5	683	PASS
174	95	50	100	96.2	46851	PASS
175	174	5	9	8.1	3799	PASS
176	174	95	101	97.2	45560	PASS
177	176	5	9	6.3	2884	PASS