

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
 Data File : VD076371.D
 Acq On : 08 Jun 2023 14:00
 Operator : KP/SY
 Sample : VD0608SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0608SBL01

Quant Time: Jun 08 16:15:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	189103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	344967	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	309862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	134071	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	97283	57.416	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.840%
35) Dibromofluoromethane	7.804	113	113583	51.857	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.720%
50) Toluene-d8	10.269	98	422309	48.426	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.860%
62) 4-Bromofluorobenzene	12.575	95	101503	43.981	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
Data File : VD076371.D
Acq On : 08 Jun 2023 14:00
Operator : KP/SY
Sample : VD0608SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0608SBL01

Quant Time: Jun 08 16:15:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

