

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
 Data File : VD076380.D
 Acq On : 08 Jun 2023 18:31
 Operator : KP/SY
 Sample : 03074-02
 Misc : 6.98G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GT-02

Quant Time: Jun 09 01:24:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	225303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	431941	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	357207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	109395	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	157507	78.024	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	156.040%
35) Dibromofluoromethane	7.799	113	208487	76.019	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	152.040%#
50) Toluene-d8	10.269	98	482995	44.233	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.460%
62) 4-Bromofluorobenzene	12.575	95	130729	45.239	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
Data File : VD076380.D
Acq On : 08 Jun 2023 18:31
Operator : KP/SY
Sample : 03074-02
Misc : 6.98G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GT-02

Quant Time: Jun 09 01:24:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

