

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
 Data File : VD076381.D
 Acq On : 08 Jun 2023 19:01
 Operator : KP/SY
 Sample : 03074-03
 Misc : 2.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GT-03

Quant Time: Jun 09 01:24:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	247314	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	440195	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	387023	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	143426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	160111	72.255	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	144.500%
35) Dibromofluoromethane	7.804	113	198694	71.090	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	142.180%
50) Toluene-d8	10.269	98	513998	46.189	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.380%
62) 4-Bromofluorobenzene	12.575	95	147595	50.118	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	100.240%
Target Compounds						
						Qvalue
16) Acetone	4.016	43	11312	27.871	ug/l	95
18) Methyl Acetate	4.575	43	3129	1.818	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
Data File : VD076381.D
Acq On : 08 Jun 2023 19:01
Operator : KP/SY
Sample : 03074-03
Misc : 2.33G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GT-03

Quant Time: Jun 09 01:24:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

