

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
 Data File : VD076383.D
 Acq On : 08 Jun 2023 20:01
 Operator : KP/SY
 Sample : 03074-05
 Misc : 7.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GT-05

Quant Time: Jun 09 01:25:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	254080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	452265	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	407019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	191139	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	146311	64.269	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.540%
35) Dibromofluoromethane	7.805	113	199929	69.623	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	139.240%
50) Toluene-d8	10.269	98	459041	40.150	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	80.300%
62) 4-Bromofluorobenzene	12.575	95	181167	59.876	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	119.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
Data File : VD076383.D
Acq On : 08 Jun 2023 20:01
Operator : KP/SY
Sample : 03074-05
Misc : 7.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GT-05

Quant Time: Jun 09 01:25:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

