

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
 Data File : VD076384.D
 Acq On : 08 Jun 2023 20:31
 Operator : KP/SY
 Sample : 03074-06
 Misc : 6.57G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GT-06

Quant Time: Jun 09 01:25:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	200618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	373011	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	307958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	177309	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	141340	78.630	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.260%
35) Dibromofluoromethane	7.799	113	175172	73.963	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	147.920%#
50) Toluene-d8	10.269	98	279333	29.623	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	59.240%
62) 4-Bromofluorobenzene	12.575	95	153459	61.495	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	122.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
Data File : VD076384.D
Acq On : 08 Jun 2023 20:31
Operator : KP/SY
Sample : 03074-06
Misc : 6.57G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GT-06

Quant Time: Jun 09 01:25:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

