

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
 Data File : VD076370.D
 Acq On : 08 Jun 2023 12:57
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 16:15:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.484	0.415	14.3	96	0.00
3 P	Chloromethane	0.930	0.799	14.1	92	0.00
4 C	Vinyl Chloride	1.221	1.051	13.9#	91	0.00
5 T	Bromomethane	0.792	0.680	14.1	91	0.00
6 T	Chloroethane	0.818	0.715	12.6	90	0.00
7 T	Trichlorofluoromethane	0.811	0.760	6.3	103	0.00
8 T	Diethyl Ether	0.264	0.247	6.4	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.519	0.507	2.3	105	0.00
10 T	Methyl Iodide	0.722	0.669	7.3	87	0.00
11 T	Tert butyl alcohol	0.025	0.022	12.0	89	-0.01
12 CM	1,1-Dichloroethene	0.551	0.515	6.5#	97	0.00
13 T	Acrolein	0.054	0.042	22.2	80	0.00
14 T	Allyl chloride	0.541	0.488	9.8	91	0.00
15 T	Acrylonitrile	0.106	0.101	4.7	95	0.00
16 T	Acetone	0.082	0.086	-4.9	98	0.00
17 T	Carbon Disulfide	1.836	1.680	8.5	95	0.00
18 T	Methyl Acetate	0.348	0.309	11.2	91	0.00
19 T	Methyl tert-butyl Ether	1.014	0.967	4.6	91	0.00
20 T	Methylene Chloride	0.782	0.646	17.4	94	0.00
21 T	trans-1,2-Dichloroethene	0.638	0.600	6.0	94	0.00
22 T	Diisopropyl ether	1.203	1.184	1.6	93	0.00
23 T	Vinyl Acetate	0.453	0.533	-17.7	106	0.00
24 P	1,1-Dichloroethane	0.983	0.960	2.3	99	0.00
25 T	2-Butanone	0.114	0.108	5.3	96	0.00
26 T	2,2-Dichloropropane	0.814	0.775	4.8	102	0.00
27 T	cis-1,2-Dichloroethene	0.689	0.665	3.5	95	0.00
28 T	Bromochloromethane	0.264	0.235	11.0	98	0.00
29 T	Tetrahydrofuran	0.067	0.061	9.0	88	0.00
30 C	Chloroform	1.031	1.005	2.5#	98	0.00
31 T	Cyclohexane	0.752	0.714	5.1	103	0.00
32 T	1,1,1-Trichloroethane	0.821	0.807	1.7	102	0.00
33 S	1,2-Dichloroethane-d4	0.448	0.439	2.0	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.317	0.337	-6.3	103	0.00
36 T	1,1-Dichloropropene	0.456	0.481	-5.5	102	0.00
37 T	Ethyl Acetate	0.150	0.146	2.7	92	0.00
38 T	Carbon Tetrachloride	0.400	0.420	-5.0	104	0.00
39 T	Methylcyclohexane	0.521	0.551	-5.8	102	0.00
40 TM	Benzene	1.443	1.492	-3.4	99	0.00
41 T	Methacrylonitrile	0.077	0.066	14.3	92	-0.01
42 TM	1,2-Dichloroethane	0.333	0.336	-0.9	97	0.00
43 T	Isopropyl Acetate	0.270	0.261	3.3	92	0.00
44 TM	Trichloroethene	0.403	0.407	-1.0	97	0.00
45 C	1,2-Dichloropropane	0.339	0.356	-5.0#	99	0.00
46 T	Dibromomethane	0.192	0.197	-2.6	97	0.00
47 T	Bromodichloromethane	0.449	0.475	-5.8	98	0.00
48 T	Methyl methacrylate	0.122	0.124	-1.6	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
 Data File : VD076370.D
 Acq On : 08 Jun 2023 12:57
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 16:15:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 06 02:52:23 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	96	0.00
50 S	Toluene-d8	1.264	1.391	-10.0	103	0.00
51 T	4-Methyl-2-Pentanone	0.142	0.147	-3.5	97	0.00
52 CM	Toluene	0.887	0.951	-7.2#	99	0.00
53 T	t-1,3-Dichloropropene	0.417	0.429	-2.9	97	0.00
54 T	cis-1,3-Dichloropropene	0.515	0.538	-4.5	97	0.00
55 T	1,1,2-Trichloroethane	0.271	0.275	-1.5	99	0.00
56 T	Ethyl methacrylate	0.277	0.290	-4.7	95	0.00
57 T	1,3-Dichloropropane	0.431	0.454	-5.3	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.116	0.123	-6.0	101	0.00
59 T	2-Hexanone	0.098	0.108	-10.2	102	0.00
60 T	Dibromochloromethane	0.308	0.331	-7.5	102	0.00
61 T	1,2-Dibromoethane	0.247	0.255	-3.2	97	0.00
62 S	4-Bromofluorobenzene	0.335	0.370	-10.4	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.358	0.368	-2.8	100	0.00
65 PM	Chlorobenzene	1.071	1.099	-2.6	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.360	0.376	-4.4	100	0.00
67 C	Ethyl Benzene	1.726	1.852	-7.3#	100	0.00
68 T	m/p-Xylenes	0.714	0.783	-9.7	101	0.00
69 T	o-Xylene	0.643	0.699	-8.7	100	0.00
70 T	Styrene	1.103	1.221	-10.7	100	0.00
71 P	Bromoform	0.194	0.207	-6.7	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.201	3.333	-4.1	103	0.00
74 T	N-amyl acetate	0.549	0.512	6.7	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.641	0.631	1.6	100	0.00
76 T	1,2,3-Trichloropropane	0.467	0.427	8.6	103	0.00
77 T	Bromobenzene	0.842	0.843	-0.1	98	0.00
78 T	n-propylbenzene	3.957	4.188	-5.8	104	0.00
79 T	2-Chlorotoluene	2.159	2.183	-1.1	102	0.00
80 T	1,3,5-Trimethylbenzene	2.639	2.778	-5.3	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.195	0.189	3.1	98	0.00
82 T	4-Chlorotoluene	2.298	2.314	-0.7	101	0.00
83 T	tert-Butylbenzene	2.213	2.350	-6.2	106	0.00
84 T	1,2,4-Trimethylbenzene	2.636	2.841	-7.8	104	0.00
85 T	sec-Butylbenzene	3.404	3.596	-5.6	106	0.00
86 T	p-Isopropyltoluene	2.822	3.056	-8.3	107	0.00
87 T	1,3-Dichlorobenzene	1.697	1.740	-2.5	103	0.00
88 T	1,4-Dichlorobenzene	1.718	1.725	-0.4	104	0.00
89 T	n-Butylbenzene	2.623	2.817	-7.4	109	0.00
90 T	Hexachloroethane	0.567	0.580	-2.3	109	0.00
91 T	1,2-Dichlorobenzene	1.473	1.491	-1.2	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.085	0.078	8.2	95	0.00
93 T	1,2,4-Trichlorobenzene	0.856	0.903	-5.5	104	0.00
94 T	Hexachlorobutadiene	0.419	0.443	-5.7	112	0.00
95 T	Naphthalene	1.600	1.610	-0.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060823\
Data File : VD076370.D
Acq On : 08 Jun 2023 12:57
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jun 08 16:15:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060523S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 06 02:52:23 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.757	0.795	-5.0	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6