

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062320\
 Data File : VD065867.D
 Acq On : 23 Jun 2020 18:25
 Operator : VA/SY
 Sample : L3082-09
 Misc : 5.90G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10

Quant Time: Jun 24 06:04:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060520S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 05 13:07:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	105780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	184744	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	161334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	58873	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	70829	67.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	135.52%	
35) Dibromofluoromethane	7.92	113	65900	56.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.94%	
50) Toluene-d8	10.34	98	211489	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	12.64	95	58940	39.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.88%	
Target Compounds						
16) Acetone	4.16	43	2998	15.205	ug/l #	76
20) Methylene Chloride	4.97	84	20039	16.123	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062320\
Data File : VD065867.D
Acq On : 23 Jun 2020 18:25
Operator : VA/SY
Sample : L3082-09
Misc : 5.90G/5.00ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10

Quant Time: Jun 24 06:04:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060520S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 05 13:07:30 2020
Response via : Initial Calibration

