

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062321\
 Data File : VD069540.D
 Acq On : 23 Jun 2021 12:57
 Operator : VA/SY
 Sample : M2811-01
 Misc : 6.21G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-04-062221

Quant Time: Jun 24 06:07:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	567181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1151370	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1181644	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	510867	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	422357	59.74	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.48%
35) Dibromofluoromethane	7.914	113	399842	51.37	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.74%
50) Toluene-d8	10.338	98	1516172	51.93	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.86%
62) 4-Bromofluorobenzene	12.626	95	560821	58.43	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.86%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062321\
Data File : VD069540.D
Acq On : 23 Jun 2021 12:57
Operator : VA/SY
Sample : M2811-01
Misc : 6.21G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-04-062221

Quant Time: Jun 24 06:07:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

