

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062321\
 Data File : VD069542.D
 Acq On : 23 Jun 2021 13:53
 Operator : VA/SY
 Sample : M2801-02RE
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-1ARE

Quant Time: Jun 24 06:08:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	145993	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	288166	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	336818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	146872	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	166604	91.55	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	183.10%#
35) Dibromofluoromethane	7.914	113	128372	65.90	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	131.80%
50) Toluene-d8	10.338	98	380761	52.11	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.22%
62) 4-Bromofluorobenzene	12.626	95	142794	59.45	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.90%
Target Compounds						
16) Acetone	4.162	43	20527	67.02	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062321\
Data File : VD069542.D
Acq On : 23 Jun 2021 13:53
Operator : VA/SY
Sample : M2801-02RE
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-1ARE

Quant Time: Jun 24 06:08:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

