

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062321\
 Data File : VD069554.D
 Acq On : 23 Jun 2021 19:31
 Operator : VA/SY
 Sample : M2835-02
 Misc : 6.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-PS-01-075

Quant Time: Jun 24 06:11:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	569506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	1098326	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	1121362	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	432800	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	413311	58.22	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.44%
35) Dibromofluoromethane	7.914	113	390382	52.58	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.16%
50) Toluene-d8	10.338	98	1436804	51.59	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.18%
62) 4-Bromofluorobenzene	12.626	95	538325	58.80	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	117.60%
Target Compounds						
7) Trichlorofluoromethane	3.268	101	58026	5.49	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062321\
Data File : VD069554.D
Acq On : 23 Jun 2021 19:31
Operator : VA/SY
Sample : M2835-02
Misc : 6.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-PS-01-075

Quant Time: Jun 24 06:11:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

