

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062323\
 Data File : VD076556.D
 Acq On : 23 Jun 2023 14:05
 Operator : KP/SY
 Sample : 03363-01RE
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SPRE

Quant Time: Jun 23 16:33:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	12207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	23877	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	24361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	10903	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	12866	105.197	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	210.400%#
35) Dibromofluoromethane	7.810	113	10229	72.114	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	144.220%
50) Toluene-d8	10.269	98	32741	60.863	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	121.720%
62) 4-Bromofluorobenzene	12.575	95	11263	64.513	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	129.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062323\
Data File : VD076556.D
Acq On : 23 Jun 2023 14:05
Operator : KP/SY
Sample : 03363-01RE
Misc : 5.06G/5.0ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SPRE

Quant Time: Jun 23 16:33:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

