

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062323\
 Data File : VD076565.D
 Acq On : 23 Jun 2023 18:33
 Operator : KP/SY
 Sample : 03361-06
 Misc : 12.84G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TOTAL-VOC-SB-29-GRAB-5

Quant Time: Jun 24 00:58:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	8766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	16253	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	10848	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	2812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.240	65	11543	131.428	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 262.860%#	
35) Dibromofluoromethane	7.804	113	8083	83.715	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 167.420%#	
50) Toluene-d8	10.269	98	15946	43.547	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 87.100%	
62) 4-Bromofluorobenzene	12.575	95	3075	25.875	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 51.760%	
Target Compounds						
					Qvalue	
64) Tetrachloroethene	10.810	164	2120	32.557	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062323\
Data File : VD076565.D
Acq On : 23 Jun 2023 18:33
Operator : KP/SY
Sample : 03361-06
Misc : 12.84G/5.0ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TOTAL-VOC-SB-29-GRAB-5

Quant Time: Jun 24 00:58:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

