

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062323\
 Data File : VD076566.D
 Acq On : 23 Jun 2023 19:03
 Operator : KP/SY
 Sample : 03361-11
 Misc : 12.61G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TOTAL-VOC-SB-31-36-SB-31-GRAB-5

Quant Time: Jun 24 00:59:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	94362	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	168603	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	127673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	36266	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	62936	66.569	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.140%
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.269	98	178376	46.958	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.920%
62) 4-Bromofluorobenzene	12.575	95	41409	33.589	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	67.180%
Target Compounds						
					Qvalue	
16) Acetone	4.022	43	143825	639.292	ug/l #	83
25) 2-Butanone	7.087	43	35565	102.208	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062323\
Data File : VD076566.D
Acq On : 23 Jun 2023 19:03
Operator : KP/SY
Sample : 03361-11
Misc : 12.61G/5.0ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TOTAL-VOC-SB-31-36-SB-31-GRAB-5

Quant Time: Jun 24 00:59:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

