

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
 Data File : VD073682.D
 Acq On : 24 Jun 2022 12:33
 Operator : VA/SY
 Sample : VD0624SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0624SBL01

Quant Time: Jun 25 07:19:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 24 07:36:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	182760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	322679	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	309811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	143400	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	122600	54.617	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.240%
35) Dibromofluoromethane	7.903	113	128547	51.139	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.280%
50) Toluene-d8	10.332	98	401587	44.495	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.000%
62) 4-Bromofluorobenzene	12.620	95	154487	45.637	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
Data File : VD073682.D
Acq On : 24 Jun 2022 12:33
Operator : VA/SY
Sample : VD0624SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0624SBL01

Quant Time: Jun 25 07:19:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 24 07:36:47 2022
Response via : Initial Calibration

