

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
 Data File : VD073687.D
 Acq On : 24 Jun 2022 15:00
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Jun 25 07:21:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 24 07:36:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	309355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	584416	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	552414	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	257614	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	184500	48.558	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.120%
35) Dibromofluoromethane	7.908	113	191123	41.981	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.960%
50) Toluene-d8	10.332	98	702986	43.006	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.020%
62) 4-Bromofluorobenzene	12.620	95	248133	40.473	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
Data File : VD073687.D
Acq On : 24 Jun 2022 15:00
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Jun 25 07:21:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 24 07:36:47 2022
Response via : Initial Calibration

