

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
 Data File : VD073694.D
 Acq On : 24 Jun 2022 18:25
 Operator : VA/SY
 Sample : N3477-03
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-04-062322

Quant Time: Jun 25 07:23:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 24 07:36:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	109038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	211118	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	216860	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	100170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	98620	73.639	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.280%
35) Dibromofluoromethane	7.909	113	81930	49.817	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.640%
50) Toluene-d8	10.332	98	260405	44.099	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.200%
62) 4-Bromofluorobenzene	12.620	95	93886	42.391	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.780%
Target Compounds						
16) Acetone	4.162	43	4032	11.543	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
Data File : VD073694.D
Acq On : 24 Jun 2022 18:25
Operator : VA/SY
Sample : N3477-03
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-04-062322

Quant Time: Jun 25 07:23:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 24 07:36:47 2022
Response via : Initial Calibration

