

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
 Data File : VD073695.D
 Acq On : 24 Jun 2022 18:54
 Operator : VA/SY
 Sample : N3502-01
 Misc : 5.82G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 D-46-UC

Manual Integrations
 APPROVED

Reviewed By : Mahesh Dadoda 06/28/2022
 Supervised By : Semsettin Yesilyurt 06/28/2022

Quant Time: Jun 25 07:23:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 24 07:36:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	213391	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	404016	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	383933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	169306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	152442	58.164	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 116.320%			
35) Dibromofluoromethane	7.908	113	147280	46.796	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 93.600%			
50) Toluene-d8	10.332	98	494599	43.768	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 87.540%			
62) 4-Bromofluorobenzene	12.620	95	174978	41.284	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 82.560%			
Target Compounds						Qvalue
16) Acetone	4.150	43	187615	450.936	ug/l	98
18) Methyl Acetate	4.703	43	5562m	2.736	ug/l	
25) 2-Butanone	7.203	43	79392	120.202	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
Data File : VD073695.D
Acq On : 24 Jun 2022 18:54
Operator : VA/SY
Sample : N3502-01
Misc : 5.82G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
D-46-UC

Quant Time: Jun 25 07:23:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 24 07:36:47 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/28/2022
Supervised By :Semsettin Yesilyurt 06/28/2022

