

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
 Data File : VD073696.D
 Acq On : 24 Jun 2022 19:24
 Operator : VA/SY
 Sample : N3502-02
 Misc : 4.75G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 D-46-LC

Quant Time: Jun 25 07:24:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 24 07:36:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	264178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	506554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	479670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	221705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	170650	52.594	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.180%
35) Dibromofluoromethane	7.909	113	173368	43.934	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.860%
50) Toluene-d8	10.332	98	617253	43.565	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.140%
62) 4-Bromofluorobenzene	12.620	95	217967	41.017	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.040%
Target Compounds						
						Qvalue
16) Acetone	4.156	43	74774	139.917	ug/l	96
25) 2-Butanone	7.209	43	23256	28.441	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
Data File : VD073696.D
Acq On : 24 Jun 2022 19:24
Operator : VA/SY
Sample : N3502-02
Misc : 4.75G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
D-46-LC

Quant Time: Jun 25 07:24:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 24 07:36:47 2022
Response via : Initial Calibration

