

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
 Data File : VD073697.D
 Acq On : 24 Jun 2022 19:53
 Operator : VA/SY
 Sample : N3502-03
 Misc : 5.44G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 D-42-UC

Quant Time: Jun 25 07:24:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 24 07:36:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	252856	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	482969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	454409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	203020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	166344	53.562	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.120%
35) Dibromofluoromethane	7.908	113	168612	44.816	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.640%
50) Toluene-d8	10.331	98	579346	42.887	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	85.780%
62) 4-Bromofluorobenzene	12.620	95	211274	41.699	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.400%
Target Compounds						
						Qvalue
16) Acetone	4.149	43	126066	252.356	ug/l	95
25) 2-Butanone	7.202	43	49676	63.472	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
Data File : VD073697.D
Acq On : 24 Jun 2022 19:53
Operator : VA/SY
Sample : N3502-03
Misc : 5.44G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
D-42-UC

Quant Time: Jun 25 07:24:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 24 07:36:47 2022
Response via : Initial Calibration

