



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062422\
 Data File : VD073680.D
 Acq On : 24 Jun 2022 09:46
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Title : SW846 8260
 Last Update : Fri Jun 24 07:36:47 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	26077	PASS
75	95	30	60	51.0	79160	PASS
95	95	100	100	100.0	155285	PASS
96	95	5	9	6.6	10312	PASS
173	174	0.00	2	1.4	1690	PASS
174	95	50	100	76.4	118600	PASS
175	174	5	9	8.1	9614	PASS
176	174	95	101	96.9	114888	PASS
177	176	5	9	8.4	9618	PASS