

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062822\
 Data File : VD073743.D
 Acq On : 28 Jun 2022 10:39
 Operator : VA/SY
 Sample : VD0628SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0628SBL01

Quant Time: Jun 28 16:27:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062722S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 16:20:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	257798	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	471879	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.631	117	446014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	204115	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	163475	80.210	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	160.420%
35) Dibromofluoromethane	7.908	113	154829	56.858	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.720%
50) Toluene-d8	10.331	98	556460	60.263	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	120.520%
62) 4-Bromofluorobenzene	12.620	95	194110	51.046	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062822\
Data File : VD073743.D
Acq On : 28 Jun 2022 10:39
Operator : VA/SY
Sample : VD0628SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0628SBL01

Quant Time: Jun 28 16:27:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062722S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 16:20:04 2022
Response via : Initial Calibration

