

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062822\
 Data File : VD073747.D
 Acq On : 28 Jun 2022 13:02
 Operator : VA/SY
 Sample : N3455-13
 Misc : 5.54G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-EXC-2-3-GR

Quant Time: Jun 28 16:30:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062722S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 16:20:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	265660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	482646	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	470623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	239845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	166603	79.282	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 158.560%	
35) Dibromofluoromethane	7.909	113	154781	55.573	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.140%	
50) Toluene-d8	10.332	98	597762	63.292	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 126.580%	
62) 4-Bromofluorobenzene	12.620	95	242315	62.302	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 124.600%	
Target Compounds						
					Qvalue	
16) Acetone	4.150	43	33367	79.046	ug/l	98
17) Carbon Disulfide	4.409	76	21997	6.427	ug/l #	94
25) 2-Butanone	7.203	43	10249	16.309	ug/l #	75
39) Methylcyclohexane	9.350	83	20581	5.835	ug/l	90
40) Benzene	8.344	78	42878	4.153	ug/l	96
52) Toluene	10.391	92	9212	1.327	ug/l	96
65) Chlorobenzene	11.661	112	11994	1.439	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062822\
Data File : VD073747.D
Acq On : 28 Jun 2022 13:02
Operator : VA/SY
Sample : N3455-13
Misc : 5.54G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-EXC-2-3-GR

Quant Time: Jun 28 16:30:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062722S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 16:20:04 2022
Response via : Initial Calibration

