

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062822\
 Data File : VD073749.D
 Acq On : 28 Jun 2022 14:00
 Operator : VA/SY
 Sample : N3492-02
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SOIL-PILE

Quant Time: Jun 29 01:42:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062722S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 16:20:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	254004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	429449	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	378900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.56

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	139766	69.079	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.160%
35) Dibromofluoromethane	7.909	113	142748	57.601	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.200%
50) Toluene-d8	10.332	98	500582	59.568	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	119.140%
62) 4-Bromofluorobenzene	12.502	95	140	0.040	ug/l	-0.12
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.080%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062822\
Data File : VD073749.D
Acq On : 28 Jun 2022 14:00
Operator : VA/SY
Sample : N3492-02
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SOIL-PILE

Quant Time: Jun 29 01:42:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062722S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 16:20:04 2022
Response via : Initial Calibration

