

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD070120\
 Data File : VD065912.D
 Acq On : 01 Jul 2020 18:03
 Operator : VA/SY
 Sample : L3172-14
 Misc : 5.07G/5.00ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP6-SS

Quant Time: Jul 02 08:23:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D062620S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 26 19:16:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	118653	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	212253	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	184006	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	74878	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	66390	53.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.72%	
35) Dibromofluoromethane	7.91	113	69006	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
50) Toluene-d8	10.34	98	231626	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
62) 4-Bromofluorobenzene	12.64	95	71045	43.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.88%	
Target Compounds						
8) Diethyl Ether	3.71	74	1904	3.285	ug/l	90
20) Methylene Chloride	4.97	84	16299	8.430	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD070120\
Data File : VD065912.D
Acq On : 01 Jul 2020 18:03
Operator : VA/SY
Sample : L3172-14
Misc : 5.07G/5.00ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP6-SS

Quant Time: Jul 02 08:23:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D062620S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 26 19:16:07 2020
Response via : Initial Calibration

